



Carestream X-Ray Film System

Certificate N°: BAM/ZBA/005/06

4th Revised version



BAM

Bundesanstalt für
Materialforschung
und -prüfung

Hereby it is confirmed by the BAM Certification Body, that the industrial X-ray film system with the designation

12200 Berlin, Germany
T: +49 30 8104-0
F: +49 30 8104-7 2222

**Industrex AA 400 Film,
processed in the
Industrex Single Part Developer,**

**automatic film processing with immersion time of 100 seconds,
developer temperature 26 °C**

of the manufacturer

**Carestream Health, Inc.
Rochester, New York, 14608
USA**

meets the requirements of **film system class C 5** for industrial X-ray films systems for radiographic testing according to **ISO 11699-1:2012** as well as the requirements of the **ASTM system class II** according to **ASTM E 1815-08 (2013)**.

Certificate holder according to certification contract

No. BAM-ZBA-0013-2007-Carestream Health S.A., 2nd Revised Version:

**Carestream Health Suisse SA
Centre „Les Tuillières“
Route de Cité-Ouest, 2
1196 Gland
Switzerland**

According to standard **ISO/IEC 17065:2012** (formerly: EN 45011:1998) the certification comprises a type examination with periodical surveillance of the products and quarterly surveillance of the manufacturing process data as well as a monitoring of the production site at Windsor, CO (BAM Certification System IIb).

The certificate is valid until **December 19, 2020**.

BAM test report **BAM VIII.3/6638** of May 9, 2006 and the results from current monitoring procedures (reference no. **BZS-GS/130/15**) are constituent parts of this certificate.

for Bundesanstalt für Materialforschung und -prüfung (BAM)
Unter den Eichen 87, 12205 Berlin, **2015-12-20**

Dr. R. Schmidt
BAM Certification Body



Dr. U. Ewert
BAM Assessor

Distribution list:: 1st Certificate holder

2nd BAM Certification Body

The BAM Certification Body has been accredited according to standard ISO/IEC 17065:2012 by the DAkkS (Deutsche Akkreditierungsstelle GmbH). The accreditation is valid for the scope given in certificate D-ZE-11075-21-00.

This certificate may only be published in full wording and without any additions. A revocable written consent shall be obtained from BAM beforehand for any amended reproduction or the publication of any excerpts. The German version is legally binding, except an English version is issued exclusively. Place of jurisdiction is Berlin.

CERTIFICATE